

## febra™ LC Long Cuff Nitrile Milking Gloves, Powder Free (50Pk)

**febra™ LC Milking Gloves** examination gloves offer high quality, cost effective, barrier protection against a wide range of general chemicals glove users come into contact with on a regular basis.

Manufactured in a dedicated Nitrile factory, this product is 100% Natural Rubber Latex free and is ideal for glove users who have been identified as having allergic reactions to NR Latex proteins (Type 1 reactions).

### Product Key Points

• Enhanced grip and feel - **Level 5 EN420** Dexterity Test Pass Value

• Fully textured palm & Fingers

• 100% NR Latex Free

• Extra Protection - Longer Cuff

• Rolled cuff prevents 'roll down'

### Product Conformance:

• EU MDD EN455 Parts 1 & 2.

• Manufactured in compliance to EU PPE 89/686/EEC "Protection Against Chemical Hazard"

• EN374 Parts 1, 2 & 3.

• EN420 Part 3.

• Passed an independent test to ISO 16604b:2003 "Resistance of Materials used in Protective Clothing to Penetration by Blood-Borne Pathogens" (viral test).

• EN420 Dexterity Test - Passed to the maximum Level 5.

• EN374 Part 3 Specific Chemical Test



### Product Information

**Type:** Powder free & non-sterile.

**Material:** Nitrile (Synthetic Latex) - Carbocyclated Acrylonitrile Butadiene.

**Protein Content:** 100% NR Latex free.

**Powder Content:** Below 2mg/g glove.

**Features:** Ambidextrous, Beaded Cuff, Textured palm & fingers.

**Colour:** Marine Blue.

**Packing:** 50 gloves per dispenser, 10 dispensers per carton.

### Dimensions

Glove Size	Product Code	Palm Width (mm)	Length (mm)
Small	GA54S	± 80	295
Medium	GA54M	± 95	295
Large	GA54L	± 110	295
Extra Large	GA54X	> 110	295

### Thickness

Location of Thickness Measurements		Single Wall (mm)
Finger	(at 15mm at the tip)	0.15 ± 0.03
Palm	(at centre of palm)	0.13 ± 0.03

### Physical Properties

Parameters	Before Aging	After Aging
Force at Break (N)	10.5	8.0
Watertight AQL	0.65	N/A

### Pre-Shipment Quality Inspection

Criteria	Insp Level	AQL
Dimensions	S-2	4.0
Physical Properties	S-2	4.0
1000ml Water Leak Test	G-1	1.5
Visual Inspection	S-4	4.0

### Quality Assurance:

• This product is manufactured in a plant where the quality management system has been independently assessed as conforming to ISO9001:2008 and ISO13485:2003.

